IJNRD.ORG ISSN : 2456-4184
INTERNATIONAL JOURNAL OF NOVEL RESEARCH AND DEVELOPMENT (IJNRD) | IJNRD.ORG
An International Open Access, Peer-reviewed, Refereed Journal

# A REVIEW OF UV-VISIBLE SPECTROSCOPY: TECHNIQUES AND APPLICATIONS

**Rajni Dubey*[1], Ashish Kumar[1], Bhaskar Kumar Gupta[2]**

[1]Associate Professor, School of Pharmacy and Research, People's University, Bhopal, Madhya Pradesh

[1]Undergraduate Students, School of Pharmacy and Research, People's University, Bhopal, Madhya Pradesh

[2]Principal, School of Pharmacy and Research, People's University, Bhopal, Madhya Pradesh

**ABSTRACT:** UV-visible spectroscopy is a versatile analytical method used to measure the absorbance of light in the UV and visible regions. It involves techniques such as single-beam and double-beam spectrophotometry and finds applications in quantitative and qualitative analysis of compounds, monitoring reaction kinetics, and material characterization. Recent advancements have improved sensitivity and precision, expanding its use in chemistry, biochemistry, and environmental science. Challenges include sample matrix interference and calibration needs, with future trends focusing on high-throughput analysis and integration with other methods.

**KEYWORDS:** Absorbance, Spectrophotometry, UV-Vis Spectrum, Quantitative Analysis Qualitative Analysis.

## INTRODUCTION

### OVERVIEW OF UV-VISIBLE SPECTROSCOPY

UV-visible spectroscopy is a powerful analytical technique used to measure the absorbance of light by a sample in the ultraviolet (UV) and visible (Vis) regions of the electromagnetic spectrum. This technique is based on the principle that different substances absorb light at specific wavelengths. By analyzing the absorbance at these wavelengths, one can obtain valuable information about the concentration and properties of the substance being studied.[1]

**Wavelength Range:** UV-visible spectroscopy typically covers the wavelength range from 200 nm to 800 nm. The UV region spans from 200 nm to 400 nm, while the visible region ranges from 400 nm to 800 nm.[4]

**Absorption Peaks:** The technique detects peaks in the absorbance spectrum, which correspond to specific electronic transitions in the molecules of the sample. These peaks are used to identify and quantify different compounds. [2]

## HISTORICAL BACKGROUND AND DEVELOPMENT

The development of UV-visible spectroscopy can be traced back to the early 20th century. Key milestones include:

**Early Discoveries:** The concept of absorption of light by matter was first explored in the 19th century, but the systematic study of UV-visible absorption began with the advent of quantum mechanics and electronic theory in the early 20th century.[1]

**First Spectrophotometers:** The first spectrophotometers capable of measuring UV-visible light were developed in the 1950s. These early instruments were relatively simple and manually operated.

**Advancements in Technology:** The 1970s and 1980s saw significant advancements with the introduction of double-beam spectrophotometers, improved monochromators, and electronic detectors. These developments greatly enhanced the sensitivity and accuracy of UV-visible spectroscopy.

**Modern Instruments:** Today's UV-visible spectrophotometers are highly sophisticated, incorporating computer control, advanced optics, and automated systems for data acquisition and analysis.[2]

## IMPORTANCE IN ANALYTICAL CHEMISTRY

UV-visible spectroscopy plays a crucial role in various fields of analytical chemistry due to its versatility and efficiency:

**Quantitative Analysis:** The technique is widely used for determining the concentration of analytes in solution. The Beer-Lambert Law, which relates absorbance to concentration, is fundamental to this application.

**Qualitative Analysis:** It aids in the identification of compounds by comparing their absorption spectra to known reference spectra. This is especially useful for detecting and characterizing organic compounds.

**Reaction Kinetics**: UV-visible spectroscopy can monitor the progress of chemical reactions in real time by observing changes in absorbance over time.[1]

**Pharmaceuticals and Biochemistry:** It is extensively used in drug development, quality control, and the study of biomolecules, including proteins and nucleic acids.

## PRINCIPLE

**ABSORPTION OF LIGHT:** UV-visible spectroscopy is based on the interaction between light and matter. Here's a breakdown of the fundamental principles:

**Electromagnetic Spectrum:** Light is composed of electromagnetic waves that vary in wavelength. UV-visible spectroscopy focuses on the UV (200-400 nm) and visible (400-800 nm) regions of the spectrum.

**Electronic Transitions:** When light passes through a sample, certain wavelengths are absorbed by the sample. This absorption occurs due to electronic transitions within the molecules. Specifically, electrons in the sample's atoms or molecules move from lower energy levels to higher energy levels when they absorb photons.

**$\pi \rightarrow \pi$ Transitions:** Common in conjugated organic compounds, where electrons in a π-bond system are excited to a π* anti-bonding orbital.

**$n \rightarrow \pi$ Transitions**: Involves non-bonding electrons (n) moving to π* anti-bonding orbitals, typically seen in molecules with lone pairs of electrons.

**Absorption Spectra:** The result of these interactions is an absorption spectrum, which plots absorbance versus wavelength. Peaks in the spectrum correspond to specific electronic transitions and can provide information about the molecular structure and concentration of the analyte.[3]

**BEER-LAMBERT LAW**

The Beer-Lambert Law describes the relationship between absorbance and concentration in a solution. It is expressed as:

$A=\log_{10}(I_0/I) = \varepsilon \cdot c \cdot l$

Where:

- A is the absorbance of the solution.
- $I_0$ is the intensity of the incident light.
- I is the intensity of the transmitted light.
- ε is the molar absorptivity (or molar extinction coefficient), which is a constant that indicates how strongly the substance absorbs light at a particular wavelength.
- c is the concentration of the analyte in the solution.
- l is the path length of the sample cell.

**Absorbance and Concentration:** Absorbance is directly proportional to concentration, provided that the solution is dilute and the path length remains constant.

**Molar Absorptivity:** This value is specific to each substance and varies with wavelength. It is determined experimentally and provides insight into the substance's absorbance characteristics.[4]




Fig.1 describes the excitation of an electron from its ground state to its excited state. This is the fundamental concept of molecular spectroscopy.

**ULTRAVIOLET (UV) LIGHT AND ITS CLASSIFICATION.**

**Near Ultraviolet (NUV)**: This range extends from approximately 300 to 400 nm. It's closest to visible light and includes UV-A, which is less harmful than other types of UV radiation but still can cause skin damage over time.

**Middle Ultraviolet (MUV)**: This range covers around 200 to 300 nm. It includes UV-B, which is more energetic than UV-A and can cause more damage to skin and eyes and is also responsible for sunburn.

**Far Ultraviolet (FUV)**: This range spans from about 100 to 200 nm. It's more energetic and can ionize atoms and molecules, which can be harmful. This includes UV-C, which is highly effective at killing bacteria and viruses but is absorbed by the Earth's atmosphere and doesn't reach the surface.

**Extreme Ultraviolet (EUV)**: With wavelengths shorter than 100 nm, EUV radiation is closer to X-rays. It's highly energetic and can penetrate materials deeply. This type is used in various scientific and industrial applications, including semiconductor manufacturing.[5]

According to ISO 21348:2007, the components and their corresponding wavelengths are defined in the table below.

| Ultraviolet Segment | Wavelength Region (10 nm - 400 nm) |
|---|---|
| Extreme Ultraviolet EUV | 10 nm - 121 nm |
| Far Ultraviolet FUV | 122 nm - 200 nm |
| Middle Ultraviolet MUV | 200 nm - 300 nm |
| Near Ultraviolet NUV | 300 nm - 400 nm |

There are two types of absorbance instruments used to collect UV-visible spectra:

1. Single beam spectrometer
2. Double beam spectrometer

**SINGLE BEAM SPECTROMETER: -**

A single-beam spectrometer is an analytical instrument utilized to measure the intensity of light across a spectrum of wavelengths. It is fundamental in fields such as chemistry, biology, and materials science for analyzing the absorbance and transmittance properties of samples.

**Components of a Single-Beam Spectrometer**

The key components of a single-beam spectrometer include a light source, monochromator, sample holder, detector, and readout device. The light source is responsible for providing a continuous spectrum of light, which can range from ultraviolet to visible light. Common light sources include deuterium lamps for UV light (160-375 nm) and tungsten lamps for visible light (350-2500 nm). Stability in the lamp's voltage is crucial to ensure a consistent energy output.[5]

The monochromator plays a vital role in isolating specific wavelengths of light from the continuous spectrum emitted by the light source. It comprises an entrance slit to narrow the light beam, a dispersive element such as a prism or diffraction grating to separate the light into its component wavelengths, and an exit slit to select a narrow band of wavelengths to pass through to the sample. The sample holder, typically a cuvette made of quartz for UV measurements or glass/plastic for visible light measurements, holds the sample in the light path. The detector measures the intensity of light after it has passed through the sample. Common detectors include photodiodes and photomultiplier tubes (PMTs), the latter being particularly sensitive to low light levels. The readout device converts the signal from the detector into a readable format, such as a digital display or computer interface, facilitating the interpretation of the measured light intensity.[6]

**Working Principle**

The operation of a single-beam spectrometer begins with the emission of light from the light source, which produces a broad spectrum of light. This light enters the monochromator, where the dispersive element separates it into its component wavelengths. The exit slit then selects the desired wavelength, which is directed towards the sample in the cuvette.

As the light passes through the sample, a portion of it is absorbed, depending on the sample's properties and concentration. The transmitted light, which is the light that has not been absorbed, reaches the detector. The detector converts the light intensity into an electrical signal, which is processed and displayed by the readout device. This process allows for the measurement of absorbance or transmittance at specific wavelengths.[7]




Fig 2. Single beam UV visible Spectrophotometer

**DOUBLE-BEAM SPECTROMETER: -** A double-beam spectrometer is an advanced analytical instrument designed to measure light intensity across a spectrum of wavelengths with high accuracy and stability. This type of spectrometer enhances measurement precision by simultaneously comparing the light passing through a sample with a reference beam. This approach compensates for any fluctuations in the light source and other environmental factors, ensuring consistent and reliable results.

**Components of a Double Beam Spectrometer**

A double-beam spectrometer consists of several key components: a light source, monochromator, beam splitter, sample and reference holders, detectors, and a readout device. The light source provides a continuous spectrum of light, which can range from ultraviolet to visible light. Typically, deuterium lamps are used for ultraviolet light (160-375 nm), and tungsten lamps are used for visible light (350-2500 nm). Maintaining a stable voltage for the light source is crucial to ensure consistent energy output. The monochromator is responsible for isolating specific wavelengths of light from the broad spectrum emitted by the light source. It consists of an entrance slit that narrows the incoming light beam, a dispersive element such as a prism or diffraction grating that separates the light into its component wavelengths, and an exit slit that selects the desired wavelength to be directed toward the sample. A beam splitter divides the monochromatic light into two beams: one directed toward the sample and the other toward the reference. This simultaneous measurement helps to compensate for any fluctuations in the light source intensity and other environmental factors. The sample holder contains the sample to be analyzed, while the reference holder contains a reference sample, often the solvent or blank. For UV measurements, quartz cuvettes are used, while glass or plastic cuvettes are suitable for visible light measurements. The detectors measure the intensity of light from both the sample and reference beams. Common detectors include photodiodes and photomultiplier tubes (PMTs), with PMTs being particularly sensitive to low light levels. The readout device converts the signals from the detectors into a readable format, such as a digital display or computer interface. This allows for precise interpretation of the measured light intensities.[8]

**Working Principle**

The operation of a double-beam spectrometer begins with the emission of light from the light source, which passes through the monochromator where the desired wavelength is selected. The monochromatic light is then split into two beams by the beam splitter. One beam passes through the sample, and the other passes through the reference. Both the sample and reference beams are directed towards their respective detectors. The sample beam passes through the sample in the cuvette, where part of the light is absorbed based on the sample's properties and concentration. The reference beam passes through the reference holder. The detectors measure the intensities of both the sample and reference beams. The readout device processes the electrical signals from the detectors, comparing the intensities of the sample and reference beams. This comparison compensates for any fluctuations in the light source intensity and other environmental factors, providing accurate and stable measurements of absorbance or transmittance.[8]




Fig 3. Double beam UV visible Spectrophotometer

**INSTRUMENTATION:** UV-visible spectroscopy (UV-Vis) is a widely used analytical technique that measures the absorption or transmission of ultraviolet and visible light by a sample. Here's an overview of the instrumentation involved in UV-Vis spectroscopy:[9]

1. Source
2. Monochromator
3. Sample cell
4. Detector
5. Computer and Software




Fig 4. Instrumentation of UV visible Spectroscopy

**1. LIGHT SOURCES**

**Deuterium Lamp Emission Range**: The deuterium lamp is used to provide continuous UV emission in the range of 160-375 nm. This range is particularly important for detecting compounds that absorb in the ultraviolet region.

**Window Material**: Quartz windows are necessary for deuterium lamps because glass absorbs UV light below 350 nm, which would interfere with accurate measurements. Quartz, on the other hand, is transparent to UV light, allowing the emitted UV radiation to pass through without significant absorption.

**Emission Range**: The tungsten-halogen lamp provides continuous visible light in the range of 350-2500 nm. This includes the entire visible spectrum and extends into the near-infrared region, making it versatile for various applications.

**Voltage Stability**: Ensuring stable lamp voltage is critical for consistent light output. Fluctuations in voltage can cause variations in light intensity, which would affect the accuracy and precision of measurements.

**2. MONOCHROMATOR**

A monochromator is a key component in UV-visible spectrophotometry that isolates specific wavelengths of light from a broader spectrum. Here's a detailed look at its structure and function:

**Function**

The primary function of a monochromator is to separate the light into its constituent wavelengths and allow only a narrow band of wavelengths to pass through to the sample. This enables precise measurement of absorbance or transmittance at specific wavelengths.

**Key Components**

**Entrance Slit**: Narrows the incoming light beam, which helps improve the spectral resolution by reducing the amount of stray light and focusing the beam onto the dispersive element.

**Design**: Typically adjustable to control the width of the light beam entering the monochromator.

1. **Dispersive Element**

**Types**:

**Diffraction Grating**: A surface with a series of closely spaced lines or grooves that diffract light into its component wavelengths. It can provide high resolution over a broad range of wavelengths.

**Prism**: A transparent optical element with a specific geometric shape that disperses light by refraction. Prisms are typically used for lower resolution but over a broader wavelength range.

**Function**: Separates the light into its constituent wavelengths by exploiting the phenomenon of diffraction or refraction.

**Exit Slit**

**Purpose**: Select a narrow band of wavelengths that have been dispersed by the dispersive element. It ensures that only the desired wavelength reaches the sample.

**Design**: Adjustable to fine-tune the width of the wavelength band passing through to the detector.

2. **Rotating Mechanism**

**Purpose**: Allows the dispersive element (usually a diffraction grating) to be rotated, thereby selecting different wavelengths to pass through the exit slit.

**Control**: The rotation is controlled by the spectrophotometer's software or manually, depending on the instrument design.

**Operation**

1. **Light Entry**: Light from the source enters through the entrance slit.
2. **Dispersion**: The dispersive element (grating or prism) separates the light into its constituent wavelengths.
3. **Wavelength Selection**: The separated light is then directed through the exit slit, which selects a specific wavelength.
4. **Transmission to Sample**: The chosen wavelength is directed towards the sample, and the intensity of the transmitted or absorbed light is measured by the detector.

**Importance**

1. **Spectral Resolution**: The monochromator's design and the width of the entrance and exit slits influence the spectral resolution of the spectrophotometer. Narrow slits and high-quality dispersive elements provide higher resolution.
2. **Wavelength Accuracy**: The precise control of the monochromator allows accurate measurement of absorbance or transmittance at specific wavelengths, essential for identifying and quantifying compounds based on their spectral characteristics.

**3. SAMPLE CELL**

1. **Cuvettes**: Containers that hold the sample. They are typically made of quartz for UV measurements and glass or plastic for visible measurements. The path length is usually 1 cm.
2. **Temperature Control**: Some instruments include temperature control for studies requiring precise temperature management.

**4. DETECTOR:**

1. **Photomultiplier Tube (PMT):** PMTs are known for their high sensitivity and low noise, making them ideal for detecting very low light levels. They work by amplifying the signal through a series of dynodes that increase the number of electrons generated from the incident photons. This makes them suitable for applications where high sensitivity is crucial, such as in low-concentration analyses or in systems with weak signals.
2. **Photodiode Array (PDA):** PDAs are composed of an array of photodiodes, each detecting light at a different wavelength simultaneously. This allows for rapid scanning and data acquisition over a range of wavelengths without moving parts. PDAs are commonly used in modern UV-Vis spectrometers for their speed and efficiency, particularly in applications requiring quick and comprehensive spectral analysis.

**5. COMPUTER AND SOFTWARE**

1. **Data Acquisition:** This system converts analog signals from the detector (like a PMT or PDA) into digital data that can be processed by a computer. The conversion process involves sampling the signal and quantifying its intensity, which is crucial for accurate measurements and analysis. High-quality data acquisition systems ensure that the data captured is both precise and representative of the actual signal.

2. **Software:** The software associated with spectrophotometers serves multiple functions. It controls the various components of the instrument, including the light source and detector. It also manages data collection, allowing users to set up experimental parameters and run analyses. Once data is collected, the software performs various calculations, such as absorbance, transmittance, and concentration, and can often generate spectral plots and perform other advanced analyses. Good software is essential for effective data interpretation and for making the most of the capabilities of the spectrophotometer.

## APPLICATIONS OF UV-VIS SPECTROSCOPY:

1. **Detection of Impurities:** UV-Vis spectroscopy can help identify impurities by measuring absorbance differences, which might indicate the presence of unintended substances.
2. **Structural Elucidation of Organic Compounds:** The spectra provide information about electronic transitions in molecules, aiding in the determination of molecular structures.
3. **Quantitative Analysis:** Used to measure the concentration of substances based on absorbance, following the Beer-Lambert law.
4. **Qualitative Analysis:** Helps in identifying compounds by analyzing their absorbance spectra and comparing them with known spectra.
5. **Chemical Analysis:** Determines the concentration and composition of chemicals in a sample.
6. **Quantitative Analysis of Pharmaceutical Substances:** Ensures accurate dosage and purity of pharmaceuticals.[10]
7. **Dissociation Constant of Acids and Bases:** Measures the strength of acids and bases by analyzing their dissociation in solution.
8. **Molecular Weight Determination:** While not the primary method, UV-Vis spectroscopy can provide indirect information related to molecular weight through structural data.
9. **As HPLC Detector:** UV-Vis detectors in HPLC measure the absorbance of compounds as they elute, providing quantitative data.
10. **Deviations from Beer-Lambert Law:** Factors like high concentrations, non-homogeneous samples, or stray light can cause deviations.

## ADVANTAGES OF UV-VIS SPECTROSCOPY:[11]

- **Speed and Ease of Use:** Provides quick and straightforward analysis.
- **Versatility:** Applicable in fields like astronomy, food safety, pharmaceuticals, and more.
- **Quality Control:** Useful in ensuring the quality and safety of products in various industries.

## DISADVANTAGES OF UV-VIS SPECTROSCOPY:[11]

- **Light Loss and Stray Light:** Can affect accuracy and sensitivity.
- **Instrument Design:** Poor design or quality of electronic components can introduce noise and reduce measurement accuracy.

## CONCLUSION:

UV-visible spectroscopy is a powerful analytical tool that enables both quantitative and qualitative analysis. It determines analyte concentration through absorbance measurements, including direct absorbance and area under the curve methods, and identifies compounds by their unique absorbance spectra. This technique is valued for its selectivity, efficiency, rapidness, and reproducibility across various scientific fields, including chemistry, biochemistry, pharmaceuticals, and environmental science. Its solid theoretical foundation and versatile applications make it indispensable for accurate and reliable analytical results.

**REFERENCES**


1. Skoog, D. A., Holler, F. J., & Crouch, S. R. (2017). *Principles of Instrumental Analysis* (7th ed.). Cengage Learning.
2. Pavia, D. L., Lampman, G. M., Kriz, G. S., & Vyvyan, J. R. (2008). *Introduction to Spectroscopy* (4th ed.). Brooks/Cole.
3. Hollas, J. M. (2004). *Modern Spectroscopy* (4th ed.). Wiley.
4. Kalshetti, B. L., Saboo, S. S., & Chaudhari, S. R. (2013). UV-Visible Spectrophotometry in Pharmaceutical Analysis. *Research Journal of Pharmacy and Technology*, 6(1), 88-93.
5. Charde, U. S., Kumar, J., & Saha, S. (2012). Applications of UV-Vis Absorption Spectroscopy in Pharmaceutical Analysis. *International Journal of Pharmaceutical Sciences and Research*, 3(1), 48-53.
6. Royal Society of Chemistry. (n.d.). UV-Visible Spectroscopy. Retrieved from https://www.rsc.org
7. Chemistry LibreTexts. (n.d.). UV-Visible Spectroscopy. Retrieved from https://chem.libretexts.org
8. Beckett A.H., Stenlake J.B., Practical Pharmaceutical Chemistry; 4th ed. CBS Publisher and Distributors Delhi, 2001: 280-286
9. Harris DC. Quantitative Chemical Analysis. 7th ed, 3rd printing. W. H. Freeman; 2007.
10. Bosch Ojeda C, Sanchez Rojas F. Recent applications in derivative ultraviolet/visible absorption spectrophotometry: 2009– 2011. Microchem J. 2013;106:1-16. doi:10.1016/j.microc.2012.05.012
11. Mend Ham J., et al; Vogel,s Textbook of Quantitative Chemical Analysis. 6th ed. Pearson Education, Singapore Pvt. Ltd ., Delhi, 2003; 6(3): 3-8.