

DEVELOPMENT OF SMART SELF-HEALING 3D-PRINTED BONE SCAFFOLD FOR EARLY FRACTURE DETECTION

¹Syed Sameer S, ²Yuvasri M.

¹Student, Department of Biomedical Engineering,

¹Rajiv Gandhi College of Engineering and Technology, Puducherry, India

²Assistant Professor, Department of Biomedical Engineering,

²Rajiv Gandhi College of Engineering and Technology, Puducherry, India

Abstract: Bone fractures, especially large skeletal defects, present a significant challenge in orthopedic healthcare. Conventional metal implants provide mechanical support but cannot monitor stress or detect early implant failure, often leading to repeated surgeries. This study focuses on developing a smart, low-cost 3D-printed bone scaffold integrated with real-time stress monitoring. The prototype uses a porous structure to mimic natural bone and embeds a flex sensor connected to an Arduino Nano microcontroller. The system processes sensor data to classify the scaffold's condition as normal, high stress, or crack risk, displays the live status on a 16x2 LCD, and triggers LED/buzzer alerts accordingly. Furthermore, a gelatin-sodium alginate hydrogel layer is incorporated to simulate the self-healing gap-filling process of natural bone. The results demonstrate that the prototype successfully detects mechanical deformation, provides immediate multi-modal feedback, and simulates the regenerative healing process. This low-cost system proves the feasibility of embedding electronics into orthopedic scaffolds, providing a strong foundation for future smart implant technologies.

Index Terms — Bone scaffold, self-healing, fracture detection, flex sensor, Arduino, 3D printing, smart implant.

INTRODUCTION

Bone is a highly dynamic living tissue with the natural ability to regenerate and heal after injury. However, when a bone defect exceeds a critical size, typically more than 2 cm, the body's natural healing process fails to bridge the gap. These critical-sized defects can occur due to severe trauma, tumor resection, or infections. Currently, the gold standard for treating such defects is the use of autografts (bone from the patient's own body) or allografts (donor bone). While effective, these methods carry significant risks, including donor site morbidity, limited availability, and the potential for immune rejection [1]. To overcome these limitations, synthetic bone scaffolds have been developed. The goal of bone tissue engineering is to create a three-dimensional structure that mimics the natural extracellular matrix, providing a temporary framework for new bone cells to attach, grow, and multiply [2]. Ideally, these scaffolds should be biodegradable, highly porous for nutrient transport, and mechanically stable. Recent advancements have utilized 3D printing to fabricate patient-specific scaffolds with precise control over pore size and geometry [3]. However, one major limitation of current orthopedic implants is their passive nature. Natural bone is "smart"; it continuously senses mechanical stress and remodels itself according to the load it experiences. In contrast, conventional metal or polymer implants cannot monitor stress, nor can they detect early signs of microfractures. As a result, implant failure is often identified only after severe structural damage has occurred, requiring complex revision surgeries [4]. To address this, researchers are exploring smart scaffolds integrated with embedded sensors. By combining biomaterials with microelectronics, it becomes possible to monitor the structural health of the implant in real time [5]. Piezoelectric materials, which generate an electrical charge in response to mechanical stress, are a popular area of study for this application [6]. Additionally, self-healing hydrogels are being developed to mimic the body's natural repair mechanism by swelling and filling micro-cracks automatically [7]. This project proposes a low-cost, proof-of-concept smart bone scaffold. Unlike complex piezoelectric systems, this prototype uses a simple flex sensor embedded within a 3D-printed bone structure. The system is controlled by an Arduino Nano and features a multi-modal alert system (LCD, LEDs, and buzzer) to warn the user of high stress or crack risk. Additionally, a gelatin-sodium alginate hydrogel layer is applied to simulate the self-healing process. The aim is to demonstrate how accessible, off-the-shelf electronics can be integrated into tissue engineering concepts to create an active monitoring system.

NEED OF THE STUDY.

The increasing rate of severe bone trauma and large skeletal defects has created a serious problem in orthopedic care. In current clinical practice, patients who receive conventional bone implants often require repeated X-rays and hospital visits to monitor bone healing. However, these imaging methods only provide static snapshots of the healing progress. They cannot detect real-time mechanical stress or identify early microfractures before they turn into catastrophic implant failures. Furthermore, the lack of real-time data means that doctors often only discover that an implant has failed after the patient experiences severe pain or loss of mobility. This leads to expensive, complicated revision surgeries that are physically and mentally draining for the patient. The knowledge of medical and paramedical staff regarding smart implant technology is still evolving, and there is a critical need for continuous, real-time monitoring systems that can bridge the gap between passive implants and active, intelligent bone scaffolds.

This study is needed to prove that low-cost, accessible electronics can be successfully embedded into a scaffold to provide early warnings, potentially preventing severe implant failure and improving patient outcomes. By integrating sensing, processing, and alert systems directly into the scaffold, this research aims to lay the foundation for future orthopedic implants that can "communicate" their health status to both the patient and the doctor.

RESEARCH METHODOLOGY

The development of the smart bone scaffold prototype was carried out through a systematic, multi-phase approach involving structural design, electronic hardware integration, and material simulation. The primary goal was to create a functional implant that could physically mimic a bone while actively monitoring its own structural health.

A. Scaffold Design and 3D Printing Parameters

The physical structure of the scaffold was designed using CAD software to closely mimic the macroscopic shape of a human long bone (such as the femur). A critical requirement for this design was the inclusion of a hollow internal cavity to securely house the microcontroller, battery pack, and wiring. This proves that the electronics can be hidden within the implant without altering its external anatomical shape. The external casing of the bone was 3D printed using Polylactic Acid (PLA) filament on an FDM printer. To mimic natural cancellous bone, the outer surface was designed with a textured, porous finish. The printing parameters were set to a layer height of 0.2 mm and a honeycomb infill pattern of 60% to provide a balance between structural rigidity and internal porosity.

B. Sensor Working Principle and Circuit Integration

The core of the monitoring system relies on an Arduino Nano microcontroller, selected for its compact size. A 2.2-inch flex sensor was integrated to act as the mechanical stress detector. The sensor operates on the principle of variable resistance—it contains a thin layer of conductive ink that stretches when bent. As the scaffold undergoes mechanical stress, the sensor bends, and its electrical resistance increases from a baseline of approximately 10 k Ω to over 30 k Ω . To convert this resistance change into a readable voltage, the flex sensor was wired into a voltage divider circuit using a 10 k Ω fixed resistor. The output voltage is fed into the analog input pin (A0) of the Arduino Nano. The Arduino uses a 10-bit Analog-to-Digital Converter (ADC) to map this voltage to a digital value between 0 and 1023. For user feedback, a 16x2 LCD was connected to digital pins 7 through 12 to show the live structural health status. The alert system consists of three LEDs (Green, Yellow, and Red) connected to digital pins 2, 3, and 4, along with a piezo buzzer on pin 5. An HC-05 Bluetooth module was also wired to the Arduino's TX/RX pins for wireless data transmission. The entire circuit was powered by a compact rechargeable battery pack embedded inside the bone cavity.

C. Self-Healing Hydrogel Preparation

To simulate the biological self-healing process, a hydrogel layer was prepared. A mixture of 5g gelatin and 2g sodium alginate was dissolved in 100 mL of distilled water and stirred continuously at 60°C until a homogeneous gel formed. This biocompatible layer is designed to absorb environmental moisture and swell when the scaffold is damaged, physically filling micro-gaps and simulating the initial stages of bone regeneration.

D. Testing and Threshold Calibration

Before final assembly, the system was tested using manual mechanical bending. The analog readings were monitored using the Arduino IDE Serial Monitor to establish the threshold values. When the scaffold was flat (unstressed), the sensor value remained below 300. When moderate bending was applied, the value fluctuated between 300 and 600. When excessive bending was applied (simulating a crack), the value exceeded 600. These calibrated thresholds were then written into the Arduino code to classify the condition into Normal (Green LED), High Stress (Yellow LED), or Crack Risk (Red LED + Buzzer).

RESULTS AND DISCUSSION

The assembled prototype was tested to evaluate its structural design, sensor accuracy, and real-time alert system. The results demonstrated that the system successfully fulfilled its intended functions: structural support, continuous stress monitoring, and visual health feedback.

System Integration and Real-Time Monitoring The electronic components, including the Arduino Nano, HC-05 Bluetooth module, and the battery pack, were successfully embedded inside the hollow cavity of the 3D-printed bone scaffold. The internal wiring was carefully routed and secured to prevent any short circuits or mechanical interference. As shown in Fig. 1, this successful

integration proves that complex electronic circuitry can be safely and compactly housed within a patient-specific implant geometry without altering the external anatomical shape of the bone.




fig. 1. internal circuit assembly showing the arduino Nano, bluetooth module, and battery pack integrated inside the scaffold.

Upon powering the system, the monitoring interface initialized successfully. Under normal, unstressed conditions, the flex sensor maintained its baseline resistance, and the system correctly classified the state as safe. As displayed in Fig. 2, the 16x2 LCD screen provided clear, real-time visual feedback by displaying "SAFE - BONE OK - H100%", indicating that the scaffold was at 100% structural health. This immediate visual confirmation is a significant improvement over conventional implants, which provide no feedback at all.



fig. 2. lcd display output showing the normal "safe" condition of the scaffold.

Stress Detection and Multi-Modal Alerts

To test the alert system, mechanical bending was manually applied to the scaffold. When moderate stress was applied, the resistance of the flex sensor increased, and the Arduino successfully triggered the yellow LED. When the bending exceeded the safe threshold—simulating a crack risk—the system immediately activated the red LED and the piezo buzzer. Fig. 3 shows the multi-modal alert system integrated into the bone head. This demonstrates that the threshold-based classification algorithm worked exactly as programmed, providing both visual and audio warnings to prevent silent implant failure. Additionally, the data was successfully transmitted via the HC-05 Bluetooth module, proving that the scaffold could theoretically send live health updates to a doctor's smartphone or a cloud database for remote IoT monitoring.




fig. 3. led and buzzer alert system integrated into the scaffold.

Self-Healing Hydrogel Observation

To simulate the biological self-healing process, the gelatin-sodium alginate hydrogel was applied to the porous surface of the scaffold. A physical cut was introduced to simulate a microfracture. Over a period of 60 minutes, the hydrogel absorbed ambient moisture and visibly swelled. The expansion of the hydrogel physically bridged the gap of the simulated fracture, restoring the structural continuity of that specific zone. While this is a physical simulation rather than true cellular regeneration, it successfully demonstrates how smart biomaterials can be integrated into the scaffold to actively respond to mechanical damage.

table 1: summary of prototype test results

Text parameter	Observed system response	Status
Hardware integration	Arduino Nano, Bluetooth, and battery successfully embedded inside the hollow cavity	Sucess
Normal condition	Flex sensor reads <300; LCDs "SAFE - BONE OK - H100%".	Sucess
High stress condition	Flex sensor reads 300–600; Yellow LED activates	Sucess
Crack risk condition	Flex sensor reads >600; Red LED and Buzzer activate.	Sucess
IoT transmission	Health data transmitted successfully via HC-05 Bluetooth module.	Sucess
Self-healing simulation	Hydrogel absorbs moisture, swells, and bridges the simulated gap.	Sucess

Discussion:

The successful operation of this prototype highlights the immense potential of combining tissue engineering with embedded electronics. The flex sensor proved to be a highly effective, low-cost alternative to complex piezoelectric sensors, accurately detecting mechanical deformation and allowing the Arduino to distinguish between normal, high stress, and crack risk conditions. Compared to conventional metal implants, which are entirely passive and can cause stress shielding (where the metal bears too much load and weakens the surrounding natural bone), this smart scaffold offers a significant advantage. It not only mimics the load distribution of natural bone but also provides continuous, real-time feedback. This allows patients and doctors to monitor healing progress actively, potentially enabling early intervention before a catastrophic implant failure occurs. When compared to recent research in smart scaffolds, most current studies rely on complex, expensive piezoelectric nanomaterials or specialized fiber Bragg grating (FBG) sensors embedded during advanced manufacturing [4], [5]. While those systems are highly accurate, they are extremely expensive and difficult to fabricate. The proposed prototype, in contrast, utilizes off-the-shelf electronics (flex sensor and Arduino) that cost under ₹5,000. This makes the technology highly accessible for educational purposes and provides a proof-of-concept for low-resource clinical settings.

Limitations and Future Work:

While the proof-of-concept was successful, a few limitations were observed. The PLA casing has lower mechanical strength than natural cortical bone and is currently suitable for prototyping, not load-bearing human applications. Additionally, the flex sensor requires careful calibration to avoid false alarms from minor body movements. The hydrogel swelling is a passive process and does not yet contain living cells. Future research will focus on replacing the flex sensor with biocompatible piezoelectric materials, using medical-grade biodegradable polymers for the scaffold, and conducting *in vitro* cell culture tests to validate biological compatibility.

ACKNOWLEDGEMENT

The authors would like to express their sincere gratitude to the management of Rajiv Gandhi College of Engineering and Technology for providing the necessary facilities and support to carry out this project successfully. We express our heartfelt gratitude to Ms. M. Yuvasri, Mentor and Assistant Professor, Department of Biomedical Engineering, for her valuable guidance, continuous encouragement, constructive suggestions, and support throughout the development and completion of this project. Her guidance and motivation played a significant role in helping us successfully carry out this work. We extend our sincere thanks to the Head of the Department and all the faculty members of the Department of Biomedical Engineering for their support and cooperation throughout the project. We also acknowledge the contributions of our colleagues and everyone who directly or indirectly supported us in completing this project on Self-Healing Scaffold for Bone Regeneration with Integrated Sensors. Finally, we express our deepest gratitude to our families for their constant encouragement, patience, and support throughout this project.

REFERENCES

- [1] P. Feng, R. Zhao, W. Tang, F. Yang, H. Tian, S. Peng, H. Pan, and C. Shuai, "Structural and functional adaptive artificial bone: Materials, fabrications, and properties," *Advanced Functional Materials*, vol. 33, no. 23, p. 2214726, Jun. 2023. [Online]. Available: <https://doi.org/10.1002/adfm.202214726>
- [2] I. Erezuma, I. Lukin, M. F. Desimone, Y. S. Zhang, A. Dolatshahi-Pirouz, and G. Orive, "Progress in self-healing hydrogels and their applications in bone tissue engineering," *Biomaterials Advances*, vol. 146, p. 213274, Mar. 2023. [Online]. Available: <https://doi.org/10.1016/j.bioadv.2022.213274>
- [3] T. Eufrásio-da-Silva, I. Erezuma, A. Dolatshahi-Pirouz, and G. Orive, "Enhancing regenerative medicine with self-healing hydrogels: A solution for tissue repair and advanced cyborg-organic healthcare devices," *Biomaterials Advances*, vol. 161, p. 213869, Sep. 2024. [Online]. Available: <https://doi.org/10.1016/j.bioadv.2024.213869>
- [4] Y. Pesaran Afsharian, C. Polley, H. Seitz, and A. Riaz, "Emerging fabrication techniques for piezoelectric scaffolds in bone tissue engineering: A systematic review," *Materials Today Advances*, vol. 31, p. 100875, Aug. 2026. [Online]. Available: <https://doi.org/10.1016/j.mtadv.2026.100875>
- [5] S. M. M. Hossen, M. A. Khaleque, M. S. Lim, J. K. Kang, D. K. Kim, H. H. Lee, and Y. Y. Kim, "Biomimetic strategies for bone regeneration: Smart scaffolds and multiscale cues," *Biomimetics*, vol. 11, no. 1, p. 12, Jan. 2026. [Online]. Available: <https://doi.org/10.3390/biomimetics11010012>
- [6] K. P. Iyengar, A. D. Kariya, R. Botchu, V. K. Jain, and R. Vaishya, "Significant capabilities of SMART sensor technology and their applications for Industry 4.0 in trauma and orthopaedics," *Sensors International*, vol. 3, p. 100163, 2022. [Online]. Available: <https://doi.org/10.1016/j.sintl.2022.100163>
- [7] A. Khodabandeh, A. A. Yousefi, S. Jafarzadeh-Holagh, and E. Vasheghani-Farahani, "Fabrication of 3D microfibrinous composite polycaprolactone/hydroxyapatite scaffolds loaded with piezoelectric poly (lactic acid) nanofibers by sequential near-field and conventional electrospinning for bone tissue engineering," *Biomaterials Advances*, vol. 166, p. 214053, Jan. 2025. [Online]. Available: <https://doi.org/10.1016/j.bioadv.2024.214053>

Copyright & License:

© Authors retain the copyright of this article. This work is published under the Creative Commons Attribution 4.0 International License (CC BY 4.0), permitting unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.